CONTROLLED//PROPIN//BASIC

United States Department of Agriculture Animal and Plant Health Inspection Service Center for Veterinary Biologics P. O. Box 844 Ames, IA 50010

1. Reagent Name: Negative Guinea Pig Serum

2. Strain or Source: Not Applicable

3. Lot Number: IRP 550 (07)

4. Fill Date: April 19, 2007

5. Expiration Date: August 31, 2028

Precautions: There are no known hazards associated with the use of this reagent.

- **6. Intended Use:** IRP 550 serves as the negative control serum when quantitating the tetanus antitoxin titer of guinea pig serum by ELISA.
- 7. Instructions for Use: When testing the potency of guinea pig tetanus antitoxin by ELISA, as described in Supplemental Assay Method (SAM) 217, dilute the serum 1:4,000 in PBS/milk/Tween diluent (0.015 M phosphate buffered saline, pH 7.2, 1% nonfat dry milk, 0.05% Tween 20). The test dilution is prepared by adding 0.1 mL of well mixed IRP 550 to 9.9 mL of PBS/milk/Tween diluent. The serum is further diluted to 1:4,000 by adding 0.1 mL of the 1:100 dilution to 3.9 mL of PBS/milk/Tween diluent.

8. Test of Reagent:

Determination of Absorbance Values - Absorbance values of less than 0.05 were recorded when IRP 550 was tested by ELISA as described in **SAM 217**.

Sterility test - The serum was tested for sterility and found to be free of viable bacteria and fungi according to procedures outlined in title 9, Code of Federal Regulations (9 CFR), section 113.26.

- **9. Container Size, Type, Weight, or Volume:** Two-mL glass vials containing 1.2 mL of serum.
- 10. Storage Conditions: Store at -20°C or lower. Once the serum has been thawed store at $4^{\circ}\pm 2^{\circ}$ C.
- **11. CVB Technical Contact:** Bacteriology Section, Center for Veterinary Biologics, (515) 337-6100.

CVB-DAT-0092.06 Page 1 of 2

CONTROLLED//PROPIN//BASIC

- 12. Origin and Passage History: Not applicable.
- **13. Method of Preparation:** Serum was collected from 500-800 gram nonvaccinated female and male guinea pigs (purchased from Small Stock and Charles River Laboratories). The serum was passed through a sterile Millipore filtration unit containing a 0.22-μm membrane, dispensed into sterile vials, and stored at -20°C.

14. Other: None

Reagent orders and feedback should be sent *including phone number* to the following email address: VS.DB.CVB.Reagent.Requests@usda.gov

Reagent order forms (APHIS Form 2018) can be found on the CVB website.

CVB-DAT-0092.06 Page 2 of 2